

Calibration certificate:  
S-A-LE 252

# TESTCERTIFICATE

---

S-A-LE 252



DORSHORST 2  
7217PH HARFSEN  
NETHERLANDS

INSTRUMENT:	WEIGHING INSTRUMENT
ID NR.:	252
MODEL:	METTLER PE 3600
CAL. DATE:	31 MARCH 2017

# WEIGHING INSTRUMENT

NR. 252: METTLER PE 3600

---

## SPECIFICATIONS

---

Range: 5 – 3600g  
Readability/Accuracy: 0,01g  
Voltage: 230 V  
Measuring temperature: 20°C +/- 5°C  
Calibration weight: Mettler Toledo 200g E2

The acceptance criterion of the deviation for repeatability and accuracy is 0,1% of the used weight. This criterion is based on an estimated acceptable contribution to the total measurement error of an analysis.

---

## MEASUREMENTS

---

Measurement	Series 1 (g)	Series 2 (g)
1	199,99	200,00
2	200,00	200,00
3	200,00	200,00
4	200,00	200,00
5	200,00	200,00
<b>Average</b>	200,00	200,00
<b>Standard deviation</b>	0,0045	0
<b>Repeatability (%)</b>	<0,01	0
<b>Accuracy (%) <sup>(1)</sup></b>	<0,01	0

<sup>(1)</sup> Expressed as deviation of the true weight

Series 1 and 2 are conducted with a weight of 200g. In series 1 the weight was solely placed on the weighing platform. In series 2 a small metal object with a mass of 670,9g was placed on the platform and the tare option was used. Then the weighing of the 200g weight was conducted.

---

## CONCLUSION

---

The instruments meets the acceptance criterion.

Calibration certificate:  
S-A-LE 252

---

**AUTHORISATION: 31-3-2017**

---



H. Kerkdijk,  
Director S-A-LE  
Date: 31-3-2017