

System: none

Instrument: ACQ-SM#L09UPA425M

Sample Manager: Back Pressure Regulator Test

Test: Passed

Start Pressure: 139 psi

End Pressure: 136 psi

Average Pressure: 137 psi

Pressure Standard Deviation: 0 psi