

Instrument: ACQ-SM#C13UPA368M

Sample Manager: Back Pressure Regulator Test

Test: Passed

Start Pressure: 141 psi

End Pressure: 140 psi

Average Pressure: 140 psi

Pressure Standard Deviation: 0 psi