

Instrument: ACQ-BSM#K09UPB811M

# ACQUITY Binary Solvent Manager Dynamic Leak Test

## Accumulator:

Result: Passed  
Leak Rate: 75.1 nL/min  
Maximum Pressure: 14032 psi  
Percent of Final Stroke: 28 %  
Compressed Volume: 16 uL  
Compression Strokes: 1

## Primary:

Result: Passed  
Leak Rate: 32.7 nL/min  
Maximum Pressure: 12023 psi  
Percent of Final Stroke: 31 %  
Compressed Volume: 18 uL  
Compression Strokes: 1

Test Attempts: 3

