

Instrument: ACQ-BSM#K09UPB811M

ACQUITY Binary Solvent Manager Dynamic Leak Test

Accumulator:

Result: Passed
Leak Rate: 29.0 nL/min
Maximum Pressure: 14025 psi
Percent of Final Stroke: 27 %
Compressed Volume: 15 uL
Compression Strokes: 1

Primary:

Result: Passed
Leak Rate: 25.0 nL/min
Maximum Pressure: 12019 psi
Percent of Final Stroke: 25 %
Compressed Volume: 14 uL
Compression Strokes: 1

Test Attempts: 1

