

System: none
Instrument: ACQ-BSM#L11UPB583A

ACQUITY Binary Solvent Manager Dynamic Leak Test

Accumulator:

Result: Passed
Leak Rate: 58.0 nL/min
Maximum Pressure: 14068 psi
Percent of Final Stroke: 27 %
Compressed Volume: 15 uL
Compression Strokes: 1

Primary:

Result: Passed
Leak Rate: 61.0 nL/min
Maximum Pressure: 12017 psi
Percent of Final Stroke: 34 %
Compressed Volume: 20 uL
Compression Strokes: 1

Test Attempts: 3

