

System: none
Instrument: ACQ-BSM#L11UPB583A

ACQUITY Binary Solvent Manager Dynamic Leak Test

Accumulator:

Result: Passed
Leak Rate: 111.0 nL/min
Maximum Pressure: 14037 psi
Percent of Final Stroke: 24 %
Compressed Volume: 13 uL
Compression Strokes: 1

Primary:

Result: Passed
Leak Rate: 81.0 nL/min
Maximum Pressure: 12021 psi
Percent of Final Stroke: 31 %
Compressed Volume: 17 uL
Compression Strokes: 1

Test Attempts: 1

