

Instrument: ACQ-SM#C13UPA368M

Sample Manager: Needle Seal Leak Test

Test: Passed

Compressed Volume: 47.9  $\mu$ L

Decompressed Volume: 32.0  $\mu$ L

Start Pressure: 107 psi

End Pressure: 113 psi

Pressure Change: 5 psi

Pressure Leak Rate: 0 psi/sec