

System: none

Instrument: ACQ-SM#L09UPA425M

Sample Manager: Needle Seal Leak Test

Test: Passed

Compressed Volume: 69.3 μ L

Decompressed Volume: 28.0 μ L

Start Pressure: 106 psi

End Pressure: 106 psi

Pressure Change: 0 psi

Pressure Leak Rate: 0 psi/sec