

Instrument: ACQ-SM#C13UPA368M

Sample Manager: Sample Syringe Leak Test

Pressure Test:

Test: Passed

Test pressure: 150 psi

Compressed Volume: 1.3 μ L

Decompressed Volume: 1.1 μ L

Start pressure: 151 psi

End pressure: 147 psi

Pressure Change: -3 psi

Average Pressure Leak: 0 psi/sec