

Instrument: ACQ-SM#L09UPA425M

Sample Manager: Sample Syringe Leak Test

Pressure Test:

Test: Passed

Test pressure: 150 psi

Compressed Volume: 1.5  $\mu$ L

Decompressed Volume: 0.7  $\mu$ L

Start pressure: 152 psi

End pressure: 105 psi

Pressure Change: -47 psi

Average Pressure Leak: -3 psi/sec